FERMI NATIONAL ACCELERATOR LABORATORY

TECHNICAL SUPPORT

VERNIER MAGNET # 002

FINAL INSPECTION DATA

LS	Q
3.956MH	1.884
3.985MH	1.882

FLOW: 8.6 GPM @ 114 PSI

EST. WT. 1,040 LBS

RESISTANCES

.1346OHMS(UPPER COIL)

.13430HMS(LOWER COIL)

7.1 GPM @ 79 PSI

DRAWING # ME-301148

WORK ORDER # M87046 ITEM 111 COST CODE W.C.R

DATE COMPLETED 10/12/88